# L-Malic acid, UV method

Catalogue number: AK00011, 58 tests (manual) / 580 tests (microplate)

**Alternative Procedures** 

## **Micro-volumes formats**

This kit has been developed to work in cuvettes with a standard pathlength of 1 cm, as described in the respective "Product Brochure". However, it can be adapted for use in 96-well microplates or in auto-analysers (micro-volume formats) with minimal assay optimisation. Basically, the assay volumes for the cuvette format have to be reduced approximately 10-fold for use in microplate format or in auto-analyser format. However, when using these micro-volume formats, you must be aware that the radiation pathlength is usually smaller than the standard cuvette pathlength of 1 cm. Thus, to perform the calculation of the amount of analyte in the samples under analysis follow one of the three strategies described in the section below.

#### **Auto-analyser procedure**

This kit is appropriate for the preparation of 139.2 mL of reagent (equivalent to 584 reactions of 0.230 mL). Reagent preparation is accomplished as follows:

### Preparation of R1:

Component	Volume
Solution 1	1.0 mL
Solution 2 (after addition of 6 mL of H2O)	1.0 mL
Suspension 3 (swirl before use)	0.2 mL
Distilled water	19.0 mL
Total	21.2 mL

#### Preparation of R2:

Component		Volume
Suspension 4 (swirl before use)		0.2 mL
Distilled water		1.9 mL
	Total	2.1 mL

## **Example Procedure:**

	Volume	
R1	0.200 mL	
Sample	0.010 mL	
Allow R1 and sample to incubate for 3 min before addition of R2		
R2	0.020 mL	

Reaction time: 3 min at 25 °C or 37 °C

Wavelength: 340 nm

**Prepared reagent stability:** > 7 days when refrigerated

Calculation: endpoint

Reaction direction: increase

 $\mbox{\bf Linearity:}$  up to 13  $\mbox{\sc \mu g/mL}$  of L-Malic acid in final reaction mixture

\* If AU values are higher than 2, please dilute the sample with distilled water accordingly.

# Strategies for analyte calculation

Auto-analysers use reaction volumes of 0.150 up to 0.6 mL and pathlengths from 4 to 8 mm, which is similar to a standard 96-well microplate in which the same reaction volume would have a pathlength of 6 or 7 mm (similar assay volumes). Therefore, in both formats (96-well microplate and auto-analysers systems), the calculation of the analyte must be done by one of the three possible methods described below:

# 1. Using the pathlength conversion factor

This is the easiest method to perform the calculation of the analyte. However, it requires a microplate reader with pathlength conversion capacity, i.e., the apparatus can detect the pathlength of each well and convert the individual readings to a 1 cm pathlength (cuvette format). In the case of auto-analysers, the absorbance readings should be directly converted to a 1 cm pathlength. This will allow the calculation of the analyte content as described in the "Product Brochure", provided with the kit and available at the NZYtech website.

#### 2. Using one standard curve

In this method, it is necessary to perform a standard curve of the analyte on each microplate that contains the test samples, or in the auto-analyser, and calculate the result from the standard curve of analyte concentration vs. absorbance. The standard curve can be performed by using the control solution provided in the kit.

#### 3. Using two standard curves

The most complicated method is to perform standard curves of the analyte in both the cuvette format (i.e. with a 1 cm of radiation pathlength) and the 96-well microplate or auto-analyser formats, and use these results to obtain a mean conversion factor between the cuvette procedure values and the alternative procedure values. The standard curves can be performed by using the control solution provided in the kit.

AK0001\_IFU\_sup\_EN\_V2301